



# Certificate of Compliance

**Certificate:** 2222827

**Master Contract:** 246454

**Project:** 2222827

**Date Issued:** October 13, 2010

**Issued to:** KEM Kueppers Elektromechanik GmbH

2 LiebigstraBe  
Karlsfeld, Bayern D85757  
Germany  
Attention: Mr. Kay Stegmann

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Lino Menezes*

**Issued by:** Lino Menezes

## **PRODUCTS**

**CLASS 2252 05** - PROCESS CONTROL EQUIPMENT

**CLASS 2252 85** - PROCESS CONTROL EQUIPMENT - Certified to US Standards

Pulse Amplifier: Model VTE, VIE or PUXxEx and SRZ or SRA Helical Flow Meter Series: Model SRZ, rated 30Vdc, Max. (SELV), 120mA; 400bar (SRZ only), ambient -20°C to 60°C and -20°C to 70°C respectively. Enclosure Type 1, 4X and 13.

Notes:

1. SELV – Safety Extra Low Voltage.
2. The “x” in the Model designations may be or followed by additional letters or numbers or blanks denoting differences in SELV secondary circuits or minor mechanical differences, not affecting the integrity of safety.

## **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

ANSI/UL Std. No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements



## *Supplement to Certificate of Compliance*

**Certificate:** 2222827

**Master Contract:** 246454

*The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
2222827	October 13, 2010	Original Certification.